



# Proxcelan™ Cesium-131 Strands

Model CS-1 (Rev.2) Seeds; Model PL-1 Strands

350 Hills Street  
Richland, WA 99354 USA  
Telephone: (509) 375-1202

**For Single Use  
Rx Only  
Sterile**

**Read these instructions and cautions carefully prior to use of product to ensure correct usage as well as safety of personnel.**

Model PL-1 Strands consist of bio-absorbable spacers and Proxcelan™ seeds (welded titanium capsules containing Cesium-131) spaced within a bio-absorbable strand material. Model PL-1 Strands have been sterilized with ethylene oxide.

Please refer to the following cautions before any manipulation of the strand.

### Cautions:

- Any manipulation of the strand should be done aseptically while working behind shielding and putting as much distance as feasible between the strand and personnel. Radiation detection equipment should be available when handling the strand.
- If it is necessary to create a shorter section of the strand, use a razor sharp scalpel or blade to perpendicularly cut the bio-absorbable strand material between the seed and spacer, using forceps to hold the strand in place.
- **Use care to avoid making contact with or cutting a seed.** A damaged seed may release radioactive Cesium-131 into the area.
- A contamination survey of the area should be performed after any manipulation of the strand to ensure seeds have not been damaged. If contamination is found in the area, the seeds should be sealed in a container and a decontamination process should be followed for the area and the personnel in the area. **Do not implant seeds that have been damaged.**

PI-004 R0



# Proxcelan™ Cesium-131 Strands

Model CS-1 (Rev.2) Seeds; Model PL-1 Strands

350 Hills Street  
Richland, WA 99354 USA  
Telephone: (509) 375-1202

**For Single Use  
Rx Only  
Sterile**

**Read these instructions and cautions carefully prior to use of product to ensure correct usage as well as safety of personnel.**

Model PL-1 Strands consist of bio-absorbable spacers and Proxcelan™ seeds (welded titanium capsules containing Cesium-131) spaced within a bio-absorbable strand material. Model PL-1 Strands have been sterilized with ethylene oxide.

Please refer to the following cautions before any manipulation of the strand.

### Cautions:

- Any manipulation of the strand should be done aseptically while working behind shielding and putting as much distance as feasible between the strand and personnel. Radiation detection equipment should be available when handling the strand.
- If it is necessary to create a shorter section of the strand, use a razor sharp scalpel or blade to perpendicularly cut the bio-absorbable strand material between the seed and spacer, using forceps to hold the strand in place.
- **Use care to avoid making contact with or cutting a seed.** A damaged seed may release radioactive Cesium-131 into the area.
- A contamination survey of the area should be performed after any manipulation of the strand to ensure seeds have not been damaged. If contamination is found in the area, the seeds should be sealed in a container and a decontamination process should be followed for the area and the personnel in the area. **Do not implant seeds that have been damaged.**

PI-004 R0